



Equipment Calibration Intervals

Revision 4.0 – Updated UL internal document approvers only. No other changes

What information you are looking for (minimum requirements):

- Equipment calibration intervals. Guidance on setting these intervals are provided in Appendix A

When should you provide this information?

- Before the visit - Test instruments are to be calibrated at intervals which maintain required tolerances specified in the test standard, manufacturer's stated tolerances or tolerances defined by the equipment owner.
- During the visit – A review of historical calibration records will provide details to confirm whether the calibration interval is appropriate. See Appendix A for details.

Why this requirement is important?

- Equipment accuracy has a direct impact on test data. Reliance on stated instrument tolerances is based on stability of the instrument during the defined calibration interval.
-

UL does not endorse any vendors or products referenced herein.

UL ASSUMES NO RESPONSIBILITY FOR ANY OMISSIONS OR ERRORS OR INACCURACIES WITH RESPECT TO THIS INFORMATION. UL MAKES NO REPRESENTATION OR WARRANTY OF ANY KIND WHATSOEVER, WHETHER EXPRESS OR IMPLIED, WITH RESPECT TO THE ACCURACY, CONDITION, QUALITY, DESCRIPTION, OR SUITABILITY OF THIS INFORMATION, INCLUDING ANY WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE AND EXPRESSLY DISCLAIMS THE SAME.

Copyright © 2012 UL LLC. All rights reserved. May not be reproduced without permission. This document is controlled and has been released electronically. The version on the UL intranet is the up-to-date document. Hard copies are uncontrolled and may not be up-to-date. Users of hard copies should confirm the revision by comparing it with the electronically controlled version.



Equipment Calibration Intervals

Certificates, Approval Forms, and Other Documentation

- Certificates and other related documentation associated with testing are to be processed in the following manner:

For WTDP -

- Historical records for calibration intervals are to be retained by the client for proof of appropriate calibration cycles. UL staff is to request copies of calibration certificates and related documentation for the equipment used in testing under the *present project*.

For other DAP programs (CTDP, TCP, TPTDP, PPP, etc.) –

- Clients are to index and retain copies of certificates and related documentation for the equipment used in testing.
- In lieu of storage of paper copies of the documentation, these may be stored electronically.
- ***Retention time for the records is to be retained for 5 years from the date of the signature of the authorized signatory on the data package in which the last time the test equipment was used.***

UL does not endorse any vendors or products referenced herein.

UL ASSUMES NO RESPONSIBILITY FOR ANY OMISSIONS OR ERRORS OR INACCURACIES WITH RESPECT TO THIS INFORMATION. UL MAKES NO REPRESENTATION OR WARRANTY OF ANY KIND WHATSOEVER, WHETHER EXPRESS OR IMPLIED, WITH RESPECT TO THE ACCURACY, CONDITION, QUALITY, DESCRIPTION, OR SUITABILITY OF THIS INFORMATION, INCLUDING ANY WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE AND EXPRESSLY DISCLAIMS THE SAME.

Copyright © 2012 UL LLC. All rights reserved. May not be reproduced without permission. This document is controlled and has been released electronically. The version on the UL intranet is the up-to-date document. Hard copies are uncontrolled and may not be up-to-date. Users of hard copies should confirm the revision by comparing it with the electronically controlled version.



Equipment Calibration Intervals

APPENDIX A

- 1 All test equipment requiring calibration shall undergo an initial calibration before being put into service. Thereafter, the maximum nominal calibration interval should be as indicated below. A calibration vendor that is accredited by a Signatory of APLAC, ILAC or EU must provide the calibration whenever possible.
 - 1A Calibration cycles shall be:
 - 1B As recommended by the manufacturer of the instrument;
 - 1C One year for electrical, electronic and mechanical test equipment; or
 - 1D Three years for mechanical test equipment made of solid materials not subject to deterioration.
- 2 Calibration intervals may be extended based on the following conditions and the reasons must be documented:
 - 2A Passive electrical test equipment, such as current shunts, current transformers, potential transformers, may be extended to 3 years with good results for the initial calibration period and if not subject to severe use conditions.
 - 2B Weights may be extended to 5 years if there is a laboratory procedure that takes into account usage and has provisions for physical examination and / or intermediate checks of the weights.
 - 2C Where there is sufficient calibration data to statistically establish a trend of the test equipment to assure good measurement results for a longer period.
3. Calibration of the following types of equipment can be put on an initial calibration only (ICO) calibration cycle:
 - steel rules,
 - tape measures,

UL does not endorse any vendors or products referenced herein.

UL ASSUMES NO RESPONSIBILITY FOR ANY OMISSIONS OR ERRORS OR INACCURACIES WITH RESPECT TO THIS INFORMATION. UL MAKES NO REPRESENTATION OR WARRANTY OF ANY KIND WHATSOEVER, WHETHER EXPRESS OR IMPLIED, WITH RESPECT TO THE ACCURACY, CONDITION, QUALITY, DESCRIPTION, OR SUITABILITY OF THIS INFORMATION, INCLUDING ANY WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE AND EXPRESSLY DISCLAIMS THE SAME.

Copyright © 2012 UL LLC. All rights reserved. May not be reproduced without permission. This document is controlled and has been released electronically. The version on the UL intranet is the up-to-date document. Hard copies are uncontrolled and may not be up-to-date. Users of hard copies should confirm the revision by comparing it with the electronically controlled version.



Equipment Calibration Intervals

- weights 4.5 kg or greater calibrated to $\pm 1\%$ tolerance,
- single piece steel probes greater than or equal to 3 mm diameter with blunt ends,
- graduate cylinders, thermometers,
- steel impact balls,
- steel or plastic probes with no moving parts and sufficient structural integrity so as to not deform
- electronic time pieces, regulated by a quartz crystal or similar stable frequency source (e.g., there are oscillator devices such as silicon-based and micro-electromechanical oscillators that can be applied in timing devices.).

With ICO, the equipment undergoes calibration and is then subjected to intermediate checks to ensure that the equipment continues to meet calibration requirements. These checks shall be conducted per a defined procedure and interval. Equipment records shall include the details on the original calibration, including certificates and the like. The last and next inspection dates shall be indicated on the equipment and in equipment records.

4. Weights do not need to be calibrated if verified with a calibrated scale before each use. The verification must be documented in the datasheets with the weighing device included in the equipment list.
5. Test equipment that is delicate, subject to frequent usage or severe conditions (i.e., shock and vibration, excessive heat or humidity, or transported) shall be assigned shortened calibration intervals.
6. Infrequently used test equipment (e.g., once or twice between calibration cycles) may be assigned the status of “Calibrate Before Use” instead of a periodic calibration.

UL does not endorse any vendors or products referenced herein.

UL ASSUMES NO RESPONSIBILITY FOR ANY OMISSIONS OR ERRORS OR INACCURACIES WITH RESPECT TO THIS INFORMATION. UL MAKES NO REPRESENTATION OR WARRANTY OF ANY KIND WHATSOEVER, WHETHER EXPRESS OR IMPLIED, WITH RESPECT TO THE ACCURACY, CONDITION, QUALITY, DESCRIPTION, OR SUITABILITY OF THIS INFORMATION, INCLUDING ANY WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE AND EXPRESSLY DISCLAIMS THE SAME.

Copyright © 2012 UL LLC. All rights reserved. May not be reproduced without permission. This document is controlled and has been released electronically. The version on the UL intranet is the up-to-date document. Hard copies are uncontrolled and may not be up-to-date. Users of hard copies should confirm the revision by comparing it with the electronically controlled version.



Equipment Calibration Intervals

UL does not endorse any vendors or products referenced herein.

UL ASSUMES NO RESPONSIBILITY FOR ANY OMISSIONS OR ERRORS OR INACCURACIES WITH RESPECT TO THIS INFORMATION. UL MAKES NO REPRESENTATION OR WARRANTY OF ANY KIND WHATSOEVER, WHETHER EXPRESS OR IMPLIED, WITH RESPECT TO THE ACCURACY, CONDITION, QUALITY, DESCRIPTION, OR SUITABILITY OF THIS INFORMATION, INCLUDING ANY WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE AND EXPRESSLY DISCLAIMS THE SAME.

Copyright © 2012 UL LLC. All rights reserved. May not be reproduced without permission. This document is controlled and has been released electronically. The version on the UL intranet is the up-to-date document. Hard copies are uncontrolled and may not be up-to-date. Users of hard copies should confirm the revision by comparing it with the electronically controlled version.